



## Phthalate Esters EPA Method 606

Column: **CH1 - 15m, 0.53mm, 1.50µm**

Catalog Code: **CH1-1553150**

### Conditions:

Injection: Split 250°C, 1.0µL.

Sample: 2000µg/mL each in Methanol.

Detector: FID 300°C.

Oven Temp.: 150°C, 15°C/min, 270°C.

Carrier Gas: Helium, 20mL/min.

### Peaks:

1. Dimethyl phthalate
2. Diethyl phthalate
3. Di-n-butyl phthalate
4. Butyl benzyl phthalate
5. Bis (2-ethylhexyl) phthalate
6. Di-n-octyl phthalate

For more information, please contact:

[info@chrom4.com](mailto:info@chrom4.com) | [www.chrom4.com](http://www.chrom4.com)

